



Certificate of Compliance

Certificate: 1376800 (LR 35072)

Master Contract: 169502

Project: 1653530

Date Issued: 2005/04/04

Issued to: **Beckman Coulter Inc.**
4300 North Harbour Blvd
Fullerton, California 92834-3100
USA
Attention: Mr. Jeff Gonzalez

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: John Chiang, P. Eng.

Authorized by: Ken Rutledge, Product Group Manager

PRODUCTS

CLASS 8721 04 - LABORATORY EQUIPMENT - Electrical

CLASS 8721 84 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

AcT Series, Models AcT 8, AcT 10, AcT Diff, rated 240/120 V ac, 50/60 Hz, 0.5/1.0 A, 120 W.

Note: The above model is Equipment Class I, Pollution Degree 2, Installation Category II.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1376800 (LR 35072)

Master Contract: 169502

Project: 1653530

Date Issued: 2005/04/04

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)

CAN/CSA-C22.2 No. 1010.1B-97 - Amendment 2 to CAN/CSA-C22.2 No. 1010.1-92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements"

UL Std No. 61010A-1 - Electrical Equipment for Laboratory Use; Part 1: General Requirements

IEC 61010-1 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Including Amendment 1:1992 and Amendment 2:1995)

CONDITIONS OF ACCEPTABILITY:

- The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.