



## Group Summary Report

Ammonia/Alcohol (XIX)

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M806651/L1						M806652/L2						M806653/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
<b>Ammonia</b>																		
ug/dL																		
Glutamate Dehyd																		
BCI CX 3/4/5/7/9	85.7	3.0	11.1	12.9	8	9	224.8		9.1	4.0	2	2	1104.5		19.8	1.8	4	4
BCI LX Systems	86.0	4.1	8.3	9.7	47	42	225.0		8.6	3.8	5	5	1114.8	12.2	15.8	1.4	27	24
BCI DxC	86.2	3.6	8.5	9.8	137	100	235.9	9.6	8.7	3.7	34	25	1110.1	10.7	17.9	1.6	91	67
All Instruments	86.1	3.7	8.5	9.9	192	151	234.0	11.6	8.7	3.7	41	32	1111.0	11.5	17.5	1.6	122	95
All Method Principles	86.1	3.7	8.5	9.9	192	151	234.0	11.6	8.7	3.7	41	32	1111.0	11.5	17.5	1.6	122	95
<b>Ammonia</b>																		
umol/L																		
Glutamate Dehyd																		
BCI CX 3/4/5/7/9	50.3	1.8	6.5	12.9	8	9	132.0		5.3	4.0	2	2	648.3		11.6	1.8	4	4
BCI LX Systems	50.5	2.4	4.9	9.7	47	42	132.1		5.0	3.8	5	5	654.4	7.2	9.3	1.4	27	24
BCI DxC	50.6	2.1	5.0	9.8	137	100	138.5	5.6	5.1	3.7	34	25	651.6	6.3	10.5	1.6	91	67
All Instruments	50.6	2.2	5.0	9.9	192	151	137.4	6.8	5.1	3.7	41	32	652.1	6.7	10.3	1.6	122	95
All Method Principles	50.6	2.2	5.0	9.9	192	151	137.4	6.8	5.1	3.7	41	32	652.1	6.7	10.3	1.6	122	95
<b>Ethanol/Alcohol</b>																		
g/L																		
Enzymatic Oxidation																		
BCI LX Systems	0.54	0.02	0.03	4.9	48	42	1.06		0.04	4.2	5	5	4.71	0.09	0.14	3.0	33	28
All Instruments	0.53	0.02	0.03	5.0	206	159	1.04	0.02	0.04	4.0	38	28	4.68	0.09	0.15	3.3	138	104
All Method Principles	0.53	0.02	0.03	5.0	206	159	1.04	0.02	0.04	4.0	38	28	4.68	0.09	0.15	3.3	138	104
<b>Ethanol/Alcohol</b>																		
g/dL																		
Enzymatic Oxidation																		
BCI LX Systems	0.054	0.002	0.003	4.9	48	42	0.106		0.004	4.2	5	5	0.471	0.009	0.014	3.0	33	28
BCI DxC	0.053	0.002	0.003	5.0	146	105	0.104	0.002	0.004	3.9	32	22	0.466	0.009	0.016	3.4	98	69
All Instruments	0.053	0.002	0.003	5.0	206	159	0.104	0.002	0.004	4.0	38	28	0.468	0.009	0.015	3.3	138	104
All Method Principles	0.053	0.002	0.003	5.0	206	159	0.104	0.002	0.004	4.0	38	28	0.468	0.009	0.015	3.3	138	104



## Group Summary Report

Ammonia/Alcohol (XIX)

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M806651/L1						M806652/L2						M806653/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
Ethanol/Alcohol																		
mg/dL																		
Enzymatic Oxidation																		
BCI CX 3/4/5/7/9	52.9	1.3	2.6	4.9	12	12	101.4		4.3	4.3	1	1	468.4	8.5	11.1	2.4	7	7
BCI LX Systems	53.5	1.5	2.6	4.9	48	42	106.4		4.5	4.2	5	5	471.0	8.9	14.2	3.0	33	28
BCI DxC	53.4	1.5	2.7	5.0	146	105	103.6	2.3	4.1	3.9	32	22	466.4	8.9	16.0	3.4	98	69
All Instruments	53.4	1.5	2.7	5.0	206	159	103.9	2.4	4.1	4.0	38	28	467.6	9.0	15.4	3.3	138	104
All Method Principles	53.4	1.5	2.7	5.0	206	159	103.9	2.4	4.1	4.0	38	28	467.6	9.0	15.4	3.3	138	104
Ethanol/Alcohol																		
mmol/L																		
Enzymatic Oxidation																		
BCI DxC	11.6	0.3	0.6	5.0	146	105	22.5	0.5	0.9	3.9	32	22	101.2	1.9	3.5	3.4	98	69
All Instruments	11.6	0.3	0.6	5.0	206	159	22.6	0.5	0.9	4.0	38	28	101.5	2.0	3.3	3.3	138	104
All Method Principles	11.6	0.3	0.6	5.0	206	159	22.6	0.5	0.9	4.0	38	28	101.5	2.0	3.3	3.3	138	104