



Group Summary Report

Ammonia/Alcohol (XX)

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M810941/L1						M810942/L2						M810943/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
Ammonia																		
ug/dL																		
Glutamate Dehyd																		
BCI CX 3/4/5/7/9	64.3	4.0	11.3	17.5	7	7	232.4		11.8	5.1	6	6	1122.3		19.9	1.8	3	3
BCI LX Systems	66.5	4.3	7.8	11.8	26	24	229.9	3.0	6.9	3.0	16	16	1109.9	11.3	14.6	1.3	20	18
BCI DxC	65.8	3.6	7.5	11.5	93	70	228.6	3.9	7.3	3.2	44	35	1101.5	13.3	14.7	1.3	84	64
All Instruments	65.8	3.8	7.8	11.9	126	101	229.3	3.8	7.6	3.3	66	57	1103.6	13.5	14.8	1.3	107	85
All Method Principles	65.8	3.8	7.8	11.9	126	101	229.3	3.8	7.6	3.3	66	57	1103.6	13.5	14.8	1.3	107	85
Ammonia																		
umol/L																		
Glutamate Dehyd																		
BCI CX 3/4/5/7/9	37.8	2.4	6.6	17.5	7	7	136.4		7.0	5.1	6	6	658.8		11.7	1.8	3	3
BCI LX Systems	39.0	2.5	4.6	11.8	26	24	135.0	1.8	4.1	3.0	16	16	651.5	6.6	8.6	1.3	20	18
BCI DxC	38.6	2.1	4.4	11.5	93	70	134.2	2.3	4.3	3.2	44	35	646.6	7.8	8.6	1.3	84	64
All Instruments	38.6	2.2	4.6	11.9	126	101	134.6	2.2	4.5	3.3	66	57	647.8	7.9	8.7	1.3	107	85
All Method Principles	38.6	2.2	4.6	11.9	126	101	134.6	2.2	4.5	3.3	66	57	647.8	7.9	8.7	1.3	107	85
Ethanol/Alcohol																		
g/dL																		
Enzymatic Oxidation																		
BCI CX 3/4/5/7/9	0.052	0.001	0.002	3.9	11	11	0.102		0.003	3.0	4	4	0.441		0.015	3.4	6	6
BCI LX Systems	0.054	0.002	0.003	5.1	27	25	0.106	0.002	0.004	4.0	16	16	0.448	0.006	0.014	3.1	20	18
BCI DxC	0.053	0.002	0.003	4.9	97	73	0.105	0.003	0.004	3.8	49	39	0.443	0.008	0.014	3.2	85	66
All Instruments	0.053	0.002	0.003	4.9	135	109	0.105	0.003	0.004	3.8	69	59	0.444	0.008	0.014	3.2	111	90
All Method Principles	0.053	0.002	0.003	4.9	135	109	0.105	0.003	0.004	3.8	69	59	0.444	0.008	0.014	3.2	112	91
Ethanol/Alcohol																		
mg/dL																		
Enzymatic Oxidation																		
BCI CX 3/4/5/7/9	52.0	1.4	2.0	3.9	11	11	101.9		3.1	3.0	4	4	440.7		14.9	3.4	6	6
BCI LX Systems	53.9	1.6	2.8	5.1	27	25	105.6	2.0	4.2	4.0	16	16	448.4	6.4	14.0	3.1	20	18
BCI DxC	53.4	1.8	2.6	4.9	97	73	104.9	2.8	4.0	3.8	49	39	443.0	8.5	14.1	3.2	85	66
All Instruments	53.4	1.8	2.6	4.9	135	109	104.9	2.6	4.0	3.8	69	59	443.9	8.3	14.1	3.2	111	90
Ion Sel(Undil-Direct)																		



Group Summary Report

Ammonia/Alcohol (XX)

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M810941/L1						M810942/L2						M810943/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
BCI LX Systems													457.8		29.9	6.5	1	1
All Instruments													457.8		29.9	6.5	1	1
All Method Principles	53.4	1.8	2.6	4.9	135	109	104.9	2.6	4.0	3.8	69	59	444.0	8.3	14.2	3.2	112	91
Ethanol/Alcohol																		
mmol/L																		
Enzymatic Oxidation																		
BCI DxC	11.6	0.4	0.6	4.9	97	73	22.8	0.6	0.9	3.8	49	39	96.2	1.8	3.1	3.2	85	66
All Instruments	11.6	0.4	0.6	4.9	135	109	22.8	0.6	0.9	3.8	69	59	96.4	1.8	3.1	3.2	111	90
All Method Principles	11.6	0.4	0.6	4.9	135	109	22.8	0.6	0.9	3.8	69	59	96.4	1.8	3.1	3.2	112	91