



# Group Summary Report

Vigil HbA1c (IX)

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M806201/L1						M806202/L2										
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS					
<b>%HbA1c (NGSP)</b>																	
percent																	
BCI CX 3/4/5/7/9	5.79		0.25	4.3	2	2	10.17	0.62	6.1		2	2	0.00	0.00	0.0	0	0
All Instruments	5.79		0.25	4.3	2	2	10.17	0.62	6.1		2	2	0.00	0.00	0.0	0	0
BCI LX Systems																	
All Instruments	5.53		0.37	6.7	1	1	9.45	0.69	7.3		1	1	0.00	0.00	0.0	0	0
All Instruments	5.53		0.37	6.7	1	1	9.45	0.69	7.3		1	1	0.00	0.00	0.0	0	0
<b>HbA1c</b>																	
g/dL																	
BCI CX 3/4/5/7/9	0.67	0.03	0.05	6.8	30	30	1.47	0.05	0.09	6.0	30	30	0.00	0.00	0.0	0	0
All Instruments	0.67	0.03	0.05	6.8	30	30	1.47	0.05	0.09	6.0	30	30	0.00	0.00	0.0	0	0
BCI LX Systems																	
All Instruments	0.67	0.04	0.04	6.0	8	8	1.46	0.05	0.07	5.0	8	8	0.00	0.00	0.0	0	0
BCI DxC	0.67	0.04	0.04	6.0	8	8	1.46	0.05	0.07	5.0	8	8	0.00	0.00	0.0	0	0
All Instruments	0.66	0.02	0.04	6.2	60	56	1.45	0.05	0.08	5.3	60	56	0.00	0.00	0.0	0	0
All Instruments	0.66	0.02	0.04	6.2	60	56	1.45	0.05	0.08	5.3	60	56	0.00	0.00	0.0	0	0
<b>Hemoglobin, Total</b>																	
g/dL																	
BCI CX 3/4/5/7/9	17.06	0.31	0.49	2.9	30	30	16.99	0.28	0.52	3.0	30	30	0.00	0.00	0.0	0	0
All Instruments	17.06	0.31	0.49	2.9	30	30	16.99	0.28	0.52	3.0	30	30	0.00	0.00	0.0	0	0
BCI LX Systems																	
All Instruments	17.10	0.50	0.46	2.7	8	8	17.10	0.55	0.43	2.5	8	8	0.00	0.00	0.0	0	0
BCI DxC	17.10	0.50	0.46	2.7	8	8	17.10	0.55	0.43	2.5	8	8	0.00	0.00	0.0	0	0
All Instruments	16.91	0.28	0.46	2.7	59	55	16.85	0.29	0.46	2.7	58	54	0.00	0.00	0.0	0	0
All Instruments	16.91	0.28	0.46	2.7	59	55	16.85	0.29	0.46	2.7	58	54	0.00	0.00	0.0	0	0