



Group Summary Report

VIGIL TDM XVI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M809571/L1						M809572/L2						M809573/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
Acetaminophen																		
ug/mL																		
Enzyme Immunoassay																		
All Method Principles	17.04	0.96	1.56	9.1	14	9	69.56	1.70	2.91	4.2	11	7	140.36	4.72	4.95	3.5	14	10
Carbamazepine																		
ug/mL																		
BCI CX/LX/DX Reagents ...																		
All Reagent Manuf	3.75	0.18	0.23	6.1	15	10	9.89	0.19	0.45	4.6	12	8	14.53	0.29	0.62	4.3	15	11
All Reagent Manuf	3.75	0.18	0.23	6.1	15	10	9.89	0.19	0.45	4.6	12	8	14.53	0.29	0.62	4.3	15	11
Digoxin																		
ng/mL																		
BCI DIGN Reagent ...																		
All Reagent Manuf	0.68	0.05	0.09	12.8	18	13	1.91	0.05	0.14	7.1	16	11	2.83	0.04	0.13	4.6	19	14
All Reagent Manuf	0.68	0.05	0.09	12.8	18	13	1.91	0.05	0.14	7.1	16	11	2.83	0.04	0.13	4.6	19	14
Gentamicin																		
ug/mL																		
BCI CX/LX/DX Reagents ...																		
All Reagent Manuf	2.45	0.18	0.13	5.2	9	7	4.09	0.11	0.17	4.2	7	6	8.08	0.17	0.30	3.7	10	9
All Reagent Manuf	2.45	0.18	0.13	5.2	9	7	4.09	0.11	0.17	4.2	7	6	8.08	0.17	0.30	3.7	10	9
Phenobarbital																		
ug/mL																		
BCI CX/LX/DX Reagents ...																		
All Reagent Manuf	8.9	0.6	0.5	5.2	19	12	23.7	1.0	1.1	4.7	15	10	45.9	2.4	2.1	4.6	19	13
All Reagent Manuf	8.9	0.6	0.5	5.2	19	12	23.7	1.0	1.1	4.7	15	10	45.9	2.4	2.1	4.6	19	13
Phenytoin																		
ug/mL																		
BCI CX/LX/DX Reagents ...																		
All Reagent Manuf	5.04	0.08	0.19	3.8	19	14	14.59	0.16	0.46	3.2	17	11	29.44	0.67	1.27	4.3	21	15
All Reagent Manuf	5.04	0.08	0.19	3.8	19	14	14.59	0.16	0.46	3.2	17	11	29.44	0.67	1.27	4.3	21	15
Salicylates																		
mg/dL																		
BCI - ENDPOINT ...																		



Group Summary Report

VIGIL TDM XVI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M809571/L1						M809572/L2						M809573/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
All Reagent Manuf	7.14	0.43	0.43	6.0	13	9	15.13	0.14	0.52	3.4	9	6	29.46	0.35	0.68	2.3	12	9
Theophylline ug/mL	7.14	0.43	0.43	6.0	13	9	15.13	0.14	0.52	3.4	9	6	29.46	0.35	0.68	2.3	12	9
BCI CX/LX/DX Reagents ...																		
All Reagent Manuf	4.59	0.10	0.24	5.3	14	9	14.41	0.22	0.56	3.9	14	9	28.79	0.84	1.05	3.6	15	10
Tobramycin ug/mL	4.59	0.10	0.24	5.3	14	9	14.41	0.22	0.56	3.9	14	9	28.79	0.84	1.05	3.6	15	10
BCI CX/LX/DX Reagents ...																		
All Reagent Manuf	1.97	0.05	0.10	5.0	7	6	3.94		0.15	3.7	4	3	8.19	0.18	0.23	2.8	8	7
Valproic Acid ug/mL	1.97	0.05	0.10	5.0	7	6	3.94		0.15	3.7	4	3	8.19	0.18	0.23	2.8	8	7
BCI CX/LX/DX Reagents ...																		
All Reagent Manuf	31.38	1.73	2.89	9.2	30	24	71.01	1.35	3.34	4.7	22	16	117.01	1.84	4.86	4.2	32	26
BCI CX/LX/DX Reagents ...	31.38	1.73	2.89	9.2	30	24	71.01	1.35	3.34	4.7	22	16	117.01	1.84	4.86	4.2	32	26
All Reagent Manuf	31.38	1.73	2.89	9.2	30	24	71.01	1.35	3.34	4.7	22	16	117.01	1.84	4.86	4.2	32	26