



## Group Summary Report

ULTIMATE-D BILIRUBIN XXI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M709681/L1						M709682/L2						M709683/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
<b>Bilirubin, Direct</b>																		
mg/dL																		
Diazo-Other (J-G) w/B																		
BCI CX 3/4/5/7/9	0.908	0.032	0.055	6.0	41	41	3.919	0.066	0.095	2.4	18	18	5.667	0.092	0.131	2.3	33	33
BCI CX 7/4CE/5CE/RTS	0.814		0.053	6.5	1	1	0.000		0.000	0.0	0	0	5.649		0.098	1.7	1	1
BCI LX Systems	0.957	0.039	0.058	6.1	47	34	3.874	0.127	0.089	2.3	18	12	5.672	0.059	0.126	2.2	43	32
BCI DxC	0.948	0.045	0.066	6.9	118	94	3.925	0.082	0.101	2.6	46	35	5.694	0.068	0.134	2.4	113	87
All Instruments	0.942	0.046	0.062	6.5	207	170	3.913	0.092	0.097	2.5	82	65	5.684	0.071	0.132	2.3	190	153
Diazo-Other(J-G) wo/B																		
BCI CX 3/4/5/7/9													5.644		0.117	2.1	1	1
All Instruments													5.644		0.117	2.1	1	1
All Method Principles	0.942	0.046	0.062	6.5	207	170	3.913	0.092	0.097	2.5	82	65	5.684	0.071	0.132	2.3	191	154
<b>Bilirubin, Total</b>																		
mg/dL																		
Diazo-Other (J-G) w/B																		
BCI LX Systems																		
All Instruments																		
Diazo-Other(J-G) wo/B																		
BCI CX 3/4/5/7/9	1.64	0.08	0.11	6.8	12	12	5.98		0.14	2.3	6	6	8.70	0.16	0.24	2.7	11	11
BCI LX Systems	1.56	0.05	0.12	7.4	27	19	5.84	0.07	0.19	3.3	13	8	8.64	0.12	0.23	2.6	25	18
BCI DxC	1.55	0.06	0.12	7.6	63	49	5.85	0.10	0.19	3.3	27	19	8.65	0.12	0.26	3.0	64	48
All Instruments	1.56	0.07	0.12	7.5	102	80	5.86	0.10	0.18	3.1	46	33	8.65	0.12	0.25	2.9	100	77
Enz Bili Ox w/B																		
BCI DxC													8.65		0.12	1.4	1	1
All Instruments													8.65		0.12	1.4	1	1
All Method Principles	1.56	0.07	0.12	7.5	102	80	5.86	0.10	0.18	3.1	46	33	8.65	0.12	0.25	2.9	101	78



**Group Summary Report**  
ULTIMATE-D BILIRUBIN XXI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M709684/L4					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS
<b>Bilirubin, Direct</b>						
mg/dL						
Diazo-Other (J-G) w/B						
BCI CX 3/4/5/7/9	10.608		0.114	1.1	1	1
BCI CX 7/4CE/5CE/RTS						
BCI LX Systems	10.613		0.144	1.4	4	3
BCI DxC	10.467		0.178	1.7	6	5
All Instruments	10.533	0.125	0.160	1.5	11	9
Diazo-Other(J-G) wo/B						
BCI CX 3/4/5/7/9						
All Instruments						
All Method Principles	10.533	0.125	0.160	1.5	11	9
<b>Bilirubin, Total</b>						
mg/dL						
Diazo-Other (J-G) w/B						
BCI LX Systems	17.69		0.54	3.0	1	1
All Instruments	17.69		0.54	3.0	1	1
Diazo-Other(J-G) wo/B						
BCI CX 3/4/5/7/9	17.22	0.32	0.45	2.6	9	9
BCI LX Systems	17.34	0.22	0.39	2.3	16	12
BCI DxC	17.49	0.28	0.43	2.5	36	27
All Instruments	17.41	0.29	0.42	2.4	61	48
Enz Bili Ox w/B						
BCI DxC	17.36		0.38	2.2	3	3
All Instruments	17.36		0.38	2.2	3	3
All Method Principles	17.41	0.28	0.42	2.4	65	52