



Group Summary Report

VIGIL SEROLOGY XVI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M804251/L1						M804252/L2						M804253/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
ASO Quant																		
IU/mL																		
BCI CX/LX/DX Rgt ...	76.3	1.7	4.0	5.3	33	24	169.4	2.7	5.9	3.5	26	20	287.7	4.7	8.4	2.9	33	24
All Reagent Manuf	76.3	1.7	4.0	5.3	33	24	169.4	2.7	5.9	3.5	26	20	287.7	4.7	8.4	2.9	33	24
BCI IMMAGE ...	84.3		3.7	4.4	5	5	156.6		8.1	5.2	3	3	304.9		17.8	5.8	3	3
All Reagent Manuf	84.3		3.7	4.4	5	5	156.6		8.1	5.2	3	3	304.9		17.8	5.8	3	3
CRP																		
mg/L																		
BCI CX/LX/DX Rgt ...	11.455	0.651	0.902	7.9	104	83	46.945	1.653	1.980	4.2	74	60	69.715	2.976	3.270	4.7	87	68
All Reagent Manuf	11.455	0.651	0.902	7.9	104	83	46.945	1.653	1.980	4.2	74	60	69.715	2.976	3.270	4.7	87	68
BCI IMMAGE ...	14.000		0.474	3.4	4	4	48.815		2.392	4.9	4	5	69.906		3.430	4.9	3	4
All Reagent Manuf	14.000		0.474	3.4	4	4	48.815		2.392	4.9	4	5	69.906		3.430	4.9	3	4
BCI hsCRP ...	15.004	0.444	0.773	5.1	103	85	50.764	1.073	1.915	3.8	104	92	71.186	1.918	3.084	4.3	11	8
All Reagent Manuf	15.004	0.444	0.773	5.1	103	85	50.764	1.073	1.915	3.8	104	92	71.186	1.918	3.084	4.3	11	8
CRP																		
mg/dL																		
BCI CX/LX/DX Rgt ...	1.146	0.065	0.090	7.9	104	83	4.694	0.165	0.198	4.2	74	60	6.971	0.298	0.327	4.7	87	68
All Reagent Manuf	1.146	0.065	0.090	7.9	104	83	4.694	0.165	0.198	4.2	74	60	6.971	0.298	0.327	4.7	87	68
BCI IMMAGE ...	1.400		0.047	3.4	4	4	4.882		0.239	4.9	4	5	6.991		0.343	4.9	3	4
All Reagent Manuf	1.400		0.047	3.4	4	4	4.882		0.239	4.9	4	5	6.991		0.343	4.9	3	4
BCI hsCRP ...	1.500	0.044	0.077	5.1	103	85	5.076	0.107	0.191	3.8	104	92	7.119	0.192	0.308	4.3	11	8
All Reagent Manuf	1.500	0.044	0.077	5.1	103	85	5.076	0.107	0.191	3.8	104	92	7.119	0.192	0.308	4.3	11	8
Rheumatoid Factor																		
IU/mL																		
Rate Nephelometry																		
BCI Image	20.0		0.0	0.0	1	1	53.6	3.1	2.8	5.3	11	11	118.6	5.2	5.2	4.4	9	9
All Instruments	20.0		0.0	0.0	1	1	53.6	3.1	2.8	5.3	11	11	118.6	5.2	5.2	4.4	9	9



Group Summary Report

VIGIL SEROLOGY XVI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M804251/L1						M804252/L2						M804253/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
Turbidometric																		
BCI CX Systems							86.9		4.7	5.4	3	3	200.7		8.4	4.2	3	3
BCI LX Systems							77.5	3.3	4.2	5.4	18	18	185.3	4.7	8.6	4.6	21	20
BCI DxC							75.4	4.3	4.8	6.4	75	59	184.3	8.9	11.1	6.0	74	56
All Instruments							76.1	4.6	4.7	6.2	96	80	185.0	8.8	10.5	5.7	98	79
All Method Principles	20.0		0.0	0.0	1	1	73.8	8.2	4.5	6.1	107	91	179.5	20.4	10.0	5.6	107	88



Group Summary Report

VIGIL SEROLOGY XVI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M804250/L4					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS
ASO Quant						
IU/mL						
BCI CX/LX/DX Rgt ...	93.8	2.5	4.5	4.8	8	6
All Reagent Manuf BCI IMMAGE ...	93.8	2.5	4.5	4.8	8	6
All Reagent Manuf						
CRP						
mg/L						
BCI CX/LX/DX Rgt ...	0.748		0.062	8.2	1	1
All Reagent Manuf BCI IMMAGE ...	0.748		0.062	8.2	1	1
All Reagent Manuf BCI hsCRP ...	0.761	0.033	0.049	6.5	106	89
All Reagent Manuf	0.761	0.033	0.049	6.5	106	89
CRP						
mg/dL						
BCI CX/LX/DX Rgt ...	0.075		0.006	8.2	1	1
All Reagent Manuf BCI IMMAGE ...	0.075		0.006	8.2	1	1
All Reagent Manuf BCI hsCRP ...	0.076	0.003	0.005	6.5	106	89
All Reagent Manuf Rate Nephelometry BCI Image All Instruments Turbidometric BCI CX Systems	0.076	0.003	0.005	6.5	106	89



Group Summary Report

VIGIL SEROLOGY XVI

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M804250/L4							
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS		
BCI LX Systems BCI DxC All Instruments All Method Principles								