



Group Summary Report

VIGIL SEROLOGY XVII

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M806331/L1						M806332/L2						M806333/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
ASO Quant																		
IU/mL																		
BCI CX/LX/DX Rgt ...	87.6	1.6	3.6	4.1	12	10	203.2	4.6	6.3	3.1	18	15	378.3	5.8	8.8	2.3	14	11
All Reagent Manuf	87.6	1.6	3.6	4.1	12	10	203.2	4.6	6.3	3.1	18	15	378.3	5.8	8.8	2.3	14	11
BCI IMMAGE ...	91.0		3.5	3.9	2	2	183.3		7.0	3.8	2	2	0.0		0.0	0.0	0	0
All Reagent Manuf	91.0		3.5	3.9	2	2	183.3		7.0	3.8	2	2	0.0		0.0	0.0	0	0
CRP																		
mg/L																		
BCI CX/LX/DX Rgt ...	10.997	0.594	0.708	6.4	62	50	49.574	1.877	1.812	3.7	60	50	75.377	3.116	2.712	3.6	70	61
All Reagent Manuf	10.997	0.594	0.708	6.4	62	50	49.574	1.877	1.812	3.7	60	50	75.377	3.116	2.712	3.6	70	61
BCI IMMAGE ...	13.887		0.421	3.0	1	1	51.033		3.015	5.9	2	2	69.350		7.535	10.9	1	1
All Reagent Manuf	13.887		0.421	3.0	1	1	51.033		3.015	5.9	2	2	69.350		7.535	10.9	1	1
BCI hsCRP ...	14.564	0.472	0.736	5.1	57	52	52.794	1.184	1.842	3.5	91	78	76.109	3.828	4.175	5.5	8	6
All Reagent Manuf	14.564	0.472	0.736	5.1	57	52	52.794	1.184	1.842	3.5	91	78	76.109	3.828	4.175	5.5	8	6
CRP																		
mg/dL																		
BCI CX/LX/DX Rgt ...	1.100	0.059	0.071	6.4	62	50	4.957	0.188	0.181	3.7	60	50	7.538	0.312	0.271	3.6	70	61
All Reagent Manuf	1.100	0.059	0.071	6.4	62	50	4.957	0.188	0.181	3.7	60	50	7.538	0.312	0.271	3.6	70	61
BCI IMMAGE ...	1.389		0.042	3.0	1	1	5.103		0.302	5.9	2	2	6.935		0.754	10.9	1	1
All Reagent Manuf	1.389		0.042	3.0	1	1	5.103		0.302	5.9	2	2	6.935		0.754	10.9	1	1
BCI hsCRP ...	1.456	0.047	0.074	5.1	57	52	5.279	0.118	0.184	3.5	91	78	7.611	0.383	0.418	5.5	8	6
All Reagent Manuf	1.456	0.047	0.074	5.1	57	52	5.279	0.118	0.184	3.5	91	78	7.611	0.383	0.418	5.5	8	6
Rheumatoid Factor																		
IU/mL																		
Rate Nephelometry																		
BCI Image	20.0		0.0	0.0	1	1	44.6	2.7	2.7	6.0	8	8	130.4		7.2	5.5	4	4
All Instruments	20.0		0.0	0.0	1	1	44.6	2.7	2.7	6.0	8	8	130.4		7.2	5.5	4	4



Group Summary Report

VIGIL SEROLOGY XVII

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M806331/L1						M806332/L2						M806333/L3					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS	AVG. MEAN	SD OF MEAN	AVG. S.D.	AVG. C.V.	NO. FLS	NO. LABS
Turbidometric																		
BCI CX 3/4/5/7/9							76.9		4.5	5.8	1	1	230.3		30.5	13.3	1	1
BCI CX Systems							76.8		3.3	4.4	3	4	246.1		11.3	4.6	5	5
BCI LX Systems							67.1	2.3	3.5	5.3	10	10	225.9	6.9	11.8	5.2	12	12
BCI DxC							66.0	4.0	3.6	5.4	61	50	217.0	11.7	11.6	5.3	58	49
All Instruments							66.7	4.5	3.6	5.3	75	65	220.5	13.1	11.9	5.4	76	67
All Method Principles	20.0		0.0	0.0	1	1	64.6	7.9	3.5	5.4	83	73	216.0	23.6	11.6	5.4	80	71



Group Summary Report

VIGIL SEROLOGY XVII

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M806330/L4					
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS
ASO Quant						
IU/mL						
BCI CX/LX/DX Rgt ...	96.4		4.2	4.4	4	3
All Reagent Manuf BCI IMMAGE ...	96.4		4.2	4.4	4	3
All Reagent Manuf						
CRP						
mg/L						
BCI CX/LX/DX Rgt ...						
All Reagent Manuf BCI IMMAGE ...						
All Reagent Manuf BCI hsCRP ...	0.751	0.045	0.045	6.0	63	56
All Reagent Manuf	0.751	0.045	0.045	6.0	63	56
CRP						
mg/dL						
BCI CX/LX/DX Rgt ...						
All Reagent Manuf BCI IMMAGE ...						
All Reagent Manuf BCI hsCRP ...	0.075	0.005	0.004	6.0	63	56
All Reagent Manuf Rate Nephelometry BCI Image All Instruments Turbidometric BCI CX 3/4/5/7/9	0.075	0.005	0.004	6.0	63	56



Group Summary Report
VIGIL SEROLOGY XVII

Lot To Date through Oct 2009

CONSTITUENT UNIT OF MEASURE METHOD PRINCIPLE INSTRUMENT/SYSTEM	M806330/L4							
	AVG. MEAN	SD OF MEAN	AVG S.D.	AVG. C.V.	NO. FLS	NO. LABS		
BCI CX Systems BCI LX Systems BCI DxC All Instruments All Method Principles								